

### Hydrometer Temperatur Correction

Temp. C°	Delta G		Temp. C°	Delta G
1	- 0.002		51	0.011
2	- 0.002		52	0.011
3	- 0.002		53	0.012
4	- 0.002		54	0.012
5	- 0.002		55	0.012
6	- 0.002		56	0.013
7	- 0.002		57	0.013
8	- 0.002		58	0.014
9	- 0.002		59	0.015
10	- 0.002		60	0.015
11	- 0.001		61	0.016
12	- 0.001		62	0.016
13	- 0.001		63	0.017
14	- 0.001		64	0.017
15	- 0.001		65	0.018
16	- 0.001		66	0.018
17	- 0.001		67	0.019
18	0		68	0.019
19	0		69	0.020
20	0		70	0.021
21	0		71	0.021
22	0		72	0.022
23	0.001		73	0.022
24	0.001		74	0.023
25	0.001		75	0.024
26	0.001		76	0.024
27	0.002		77	0.025
28	0.002		78	0.025
29	0.002		79	0.026
30	0.002		80	0.027
31	0.003		81	0.027
32	0.003		82	0.028
34	0.004		83	0.029
35	0.004		84	0.029
36	0.004		85	0.030
37	0.005		86	0.031
38	0.005		87	0.031
39	0.006		88	0.032
40	0.006		89	0.033
41	0.006		90	0.033
42	0.007		91	0.034
43	0.007		92	0.035
44	0.007		93	0.035
45	0.008		94	0.036
46	0.008		95	0.037
47	0.009		96	0.037
48	0.009		97	0.038
49	0.010		98	0.039
50	0.010		99	0.039

Made By  
Søren Ladegaard

Visit  
[www.haandbrygforum.dk](http://www.haandbrygforum.dk)

**Calculations made with ProMash 1.8.a  
For use with Brouland hydrometer ( 013.007.0 )**